



UNITS ARE IN mm
UNLESS OTHERWISE
NOTED



$$Y = E_L - E$$

TOOL REFERENCE



Template# 0035

Fill Level to Base of Neck= 9ml Approx.

IMAGE NOT TO SCALE

Approvals:

Drawn:		Quality:		Manufacturing:		Customer:	
DEN		AAP		BKE			
Date 10/28/2005		Date 1/6/2006		Date 1/11/2006		Date	
A =	22.25 ± 0.25	J =	0.76 MAX.	R =		Z =	
B =	1.25 ± 0.05	K =	1.78 R MIN.	S =		AA =	
C =	39.50 ± 0.50	L =	1.45 MIN.	T =		BB =	
D =	19.95 ± 0.25	M =	3.60 ± 0.20	U =		CC =	
E =	12.70 ± 0.25	N =		V =		DD =	
F =		O =		W =		EE =	
G =		P =	7.50 ± 0.50	X =		FF =	
H =	0.76 MIN.	Q =		Y =		E _L =	
Revision	ECN Number	Date		By			
A	10524	03/12/99		CAW			
Glass Type:		USP TYPE 1		Cust.:		West Pharmaceutical Services	
Desc.:		VIA CRP 0050 FC 19.95 ABB SL 22.25 1.25		Item Number:		1144246	
SCHOTT		Customer Item Number:		6800-0318		Date Issued:	
		Ref. Item Number:		68000318		06/30/99	
		Drawing Number:		24205080			

Rik-Keene
31/Mar 2010